



Natriumthiosulfat Titrationskartusche (DT), 0,113 N

Artikel-Nr.: 2267301

CHF Preis: Kontakt

Verfügbar

Gebrauchsfertige Titrierlösung für die Digital Titrator Titration von Gesamt Chlor, 20-2000 mg/l Cl₂.

For chlorine determination by iodometric Digital Titrator titration. 0.113 N, Digits x 0.1 = mg/L Cl₂ in a 5 mL sample, Digits x 0.000141 = meq. Approximately 40-100 tests

Accurate results - when and where you need them

Digital Titrator methods use a high-precision dispensation device and cartridges of concentrated titrant to allow analyses in the lab, the plant or the field.

Fast and Flexible

Interchangeable cartridges allow you to change titrants in seconds. No cleaning or rinsing burets.

Versatile

More than 50 parameter/range combinations are available

Economical

Concentrated titrants allow 40-100 typical titrations per cartridge.

Technische Daten

Anzahl Bestimmungen:	40 - 100
Messbereich:	20 - 2000 mg/L Cl ₂
Methode:	Iodometric
Methode (Nummer):	8209
Parameter:	Chlorine, Total
Plattform:	Digital Titrator
Titrant:	0.113 N Na ₂ S ₂ O ₃