

Digital titrator cartridge, sodium thiosulfate, 0.2068N

Artikel-Nr.: **2267601**

CHF Preis: Kontakt
Kein Versanddatum angezeigt

For determination of Chromate by Digital Titrator titration. 0.2068 N Na2S2O3. Digits x 0.2 = mg/L CrO4 in a 50 mL sample.

Accurate results - when and where you need them
Digital Titrator methods use a high-precision dispensation device and cartridges of concentrated titrant to allow analyses in the lab, the plant or the field.

Fast and Flexible
Interchangeable cartridges allow you to change titrants in seconds. No cleaning or rinsing burets.

Versatile
More than 50 parameter/range combinations are available

Economical
Concentrated titrants allow 40-100 typical titrations per cartridge.

Technische Daten

Anzahl Bestimmungen:	40 - 100
Messbereich:	20 - 400 mg/L CrO_4
Methode:	Iodometric
Methode (Nummer):	8211
Parameter:	Chromate
Plattform:	Digital Titrator
Titrant:	0.2068 N $Na_2S_2O_3$